M
11573

MODEL 252

TEGAM

MODEL 252

MODEL 252

Digital Impedance Meter

Instruction Manual

Part Number 43158E February 1991



TEGAM, INC., Ten Tegam Way, Geneva, Ohio 44041 440 466-6100 Fax: 440 466-6110

PROTECT AGAINST ELECTROSTATIC DISCHARGE

In this instrument there are MOSFET semiconductors, which can be damaged by electrostatic discharge during handling. The following precautionary procedures are recommended to minimize this possibility.

DISCHARGE PERSONAL STATIC BEFORE HANDLING DEVICES

HANDLE STATIC SENSITIVE DEVICES ONLY AT A GROUNDED, STATIC-FREE WORK STATION

HANDLE DEVICES BY THE BODY. DO NOT TOUCH THE DEVICE LEADS.

BE SURE YOUR SOLDERING IRON TIP IS GROUNDED AND DO NOT USE SOLDER-SUCKERS THAT ARE NOT ANTI—STATIC PROTECTED

AVOID HANDLING WHENEVER POSSIBLE

USE ANTI-STATIC PACKAGING FOR HANDLING AND TRANSPORT

KEEP PARTS IN MANUFACTURER'S PROTECTED CONTAINERS

DO NOT SLIDE STATIC SENSITIVE DEVICES OVER ANY SURFACE AND AVOID PLASTIC, VINYL AND STYROFOAM IN WORK AREAS

TEGAM, INC., Ten Tegam Way, Geneva, Ohio 44041 440 466-6100 Fax: 440 466-6110

TABLE OF CONTENTS

WARNING

DANGEROUS VOLTAGE POTENTIALS EXIST INSIDE THIS INSTRUMENT. MAINTENANCE INSTRUCTIONS WITHIN THIS MANUAL ARE FOR USE BY QUALIFIED SERVICE PERSONNEL ONLY. TO AVOID ELECTRICAL SHOCK, DO NOT ATTEMPT ANY SERVICING OTHER THAN THAT CONTAINED IN THE OPERATION INSTRUCTIONS UNLESS YOU ARE QUALIFIED TO DO SO.

1 DESCRIPTION

1.1 INTRODUCTION 1-1
1.2 SPECIFICATIONS 1-2

2 OPERATION

2.1 CONTROLS AND CONNECTORS 2-1
2.1.1 Front Panel 2-1
2.1.2 Rear Panel 2-3

2.2 OPERATING PROCEDURE 2-4
2.2.1 Power Requirements 2-4
2.2.2 Applying Power 2-5
2.2.3 Connection to Unknowns 2-6
2.2.3.1 Test Fixture Compensation 2-8
2.2.4 Function Selection 2-9
2.2.5 Range Selection 2-12
2.2.6 Summary of Operation 2-13

2.3 MEASUREMENT TECHNIQUES 2-15
2.3.1 Resistance Measurements at 1kHz 2-15
2.3.2 Capacitance Measurements 2-15
2.3.3 Inductance Measurements 2-18
2.3.4 Determining Quality Factor (Q) of Inductors 2-19
2.3.5 Determining Dissipation Factor (D) of Capacitors using Nomograph 2-20
2.3.6 Using the Bias Feature 2-22
2.3.7 Signal Output Terminals 2-23
2.3.8 Measuring Grounded Unknowns 2-23
2.3.9 Measuring Battery Impedance 2-26
2.3.10 Component Sorting 2-28

3 CIRCUIT DESCRIPTIONS

3.1 GENERAL DESCRIPTION 3-1

TABLE OF CONTENTS (continued)

3.2 DETAILED CIRCUIT DESCRIPTIONS 3-2
3.2.1 1kHz Oscillator 3-3
3.2.2 Unknown Differential Amplifier 3-3
3.2.3 Range Amplifier 3-3
3.2.4 Reference Generator and Phase Sensitive Detector 3-4
3.2.5 Dissipation Factor Circuit 3-4
3.2.6 Overload Blanking Circuit 3-5
3.2.7 Analog to Digital Converter 3-5
3.2.8 Function and Range Programming 3-5

4 MAINTENANCE

4.1 CALIBRATION PROCEDURE 4-1
4.1.1 Power Supply Check 4-2
4.1.2 Oscillator Adjust 4-2
4.1.3 L, C, R, G Alignment 4-2
4.1.4 Dissipation Factor (D) Alignment 4-3

4.2 MAINTENANCE 4-7
4.2.1 Preventive Maintenance 4-7
4.2.1.1 Cleaning 4-7
4.2.1.2 Visual Inspection 4-8
4.2.2 Troubleshooting 4-8
4.2.2.1 Troubleshooting Aids 4-8
4.2.2.2 Troubleshooting Procedure 4-9

4.3 CORRECTIVE MAINTENANCE 4-12
4.3.1 Obtaining Replacement Parts 4-12
4.3.2 Component Replacement Parts 4-13

4.4 REPACKING FOR SHIPMENT 4-14

5 PARTS LISTS AND DIAGRAMS

5.1 MAINFRAME PARTS 5-1
5.2 MOTHERBOARD CIRCUIT ASSEMBLY 5-2
5.3 DISPLAY CIRCUIT ASSEMBLY 5-8
5.4 BACK PANEL ASSEMBLY 5-10
5.5 POWER SUPPLY CIRCUIT ASSEMBLY 5-12

SECTION 1
DESCRIPTION

1.1 INTRODUCTION

Model 252 Digital Impedance Meter is a semi-automatic instrument which permits rapid measurement of inductance (L), capacitance (C), resistance (R), conductance (G) and dissipation factor (D) at a test frequency of 1kHz. Measurement accuracy and versatility satisfies most demanding engineering or scientific applications.

To operate, merely push the button for the desired function, manually turn the knob to the desired range, and connect the unknown. KELVIN KLIPS® test leads are included, thus ensuring true four-terminal connections. The position of the range switch, used in conjunction with the desired function button and front panel range scale, indicate the unit of measurement being displayed by the 3-1/2 digit LED readout.

Excellent reliability of the Model 252 is assured through use of solid state devices and etched circuit board construction. Its small size is ideal for use on benchtops where work space may be at a premium. The carrying handle tilts the unit to a convenient viewing angle. Rear panel brackets provide line cord storage and enable it to be operated in a vertical position.

1.2 SPECIFICATIONS

		Ranges						
		0	1	2	3	4	5	6
Full Scale Values	L_S	200 µH	2 mH	20 mH	200 mH	2 H	20 H	200 H
	C_P	200 pF	2 nF	20 nF	200 nF	2 µF	20 µF	200 µF
	R_S	2 Ω	20 Ω	200 Ω	2 kΩ	20 kΩ	200 kΩ	2000 kΩ
	G_P	2 µS	20 µS	200 µS	2 mS	20 mS	200 mS	2000 mS
	D	1.999						
Accuracy (15°C to 35°C)	L_S	±[0.25% + (1 + 0.002R_S*) digits] **	±[0.25% + (1 + 0.001R_S*)digits]					±[0.25% + (1+0.010R_S*) digits]
	C_P	±[0.25% + (1 + 0.002G_P*) digits] **	±[0.25% + (1 + 0.001G_P*)digits]					±[0.25% + (1+0.010G_P*) digits]
	R_S	±[0.25% + (1 + 0.002L_S*) digits]	±[0.25% + (1 + 0.001L_S*)digits]					±[0.25% + (1+0.010L_S*) digits]
	G_P	±[0.25% + (1 + 0.002C_P*) digits]	±[0.25% + (1 + 0.001C_P*)digits]					±[0.25% + (1+0.010C_P*) digits]
	D	±(1% + 0.002) for L or C ≥ 200 counts ±(2% + 0.010) for L or C 50 to 199 counts						
Test Signal	Voltage C_P, G_P	1.0 VRMS		0.1 VRMS				
	Current L_S, R_S	100 mA	10 mA	1 mA	100 µA	10 µA		1 µA

*Digit count, same range.
**After correction for test lead zero reading.
0°C to 15°C and 35°C to 50°C: add 0.1(rated accuracy)/°C.

Table 1-1. Model 252 Specifications

Test Frequency: 1kHz ±1%.

Unknown Excitation: The 1kHz voltage (Vx) and current (Ix) levels listed in Table 1-1 are held constant by an internal amplitude control circuit.

Measurement Rate: Four per second; one second is required for first reading after connecting unknown to terminals.

Measurement Display: 3-1/2 digit LED with decimal point. Blanked for overload conditions.

Unit Display: Unit of measurement being displayed by the LED readout is indicated by position of the range switch, used in conjunction with the desired function button and the front panel range scale.

External Bias: Rear panel terminals are provided for connection of external supply. 0V to 50VDC, 0.1A maximum. (Read Section 2.2.6 before using external bias.)

Static Charge Protection: Diode and resistor discharge network.

Connection to Unknown: Four-terminal, shielded, connections are provided by the KELVIN KLIPS® cable assembly (Part Number 43072) supplied with the Model 252.

Outputs: Analog signals of 1V per 1,000 counts, 1kΩ source resistance is available at rear panel. L, C, R, or G, with simultaneous output of D for L or C.

Power Consumption: 4 watts typical.

Power Requirements: 100 to 125V or 200 to 250V, 50/60Hz.

Fuse: 110V: 1/16A 250VAC Slow-Blow
220V: 1/20A 250VAC Slow-Blow

Size: Height (with feet) - 100mm (4 in.)
Width - 260mm (10 in.)
Depth (overall) - 370mm (14.6 in.)

Weight: 3.2kg (7 lb).

Accessories supplied with Model 252:

	Part No.
KELVIN KLIPS® Four-Terminal Clips	43072
Instruction Manual	43158

Options Available:

	Part No.
Model 1412B Universal Limits Comparator	31412B
Sorting Fixture Model 2001 (low frequency)	32001
Cable Assembly (for Model 2001 connection)	43586
Front Panel Dust Cover	43374

SECTION 2
OPERATION

2.1 CONTROLS AND CONNECTORS

2.1.1 Front Panel



Figure 2-1. Model 252 Front Panel

1. ON/OFF Switch -- A push-on, push-off switch for applying and removing power from the instrument.

2. UNKNOWN Connector -- Terminals designed to be used with KELVIN KLIPS® test leads (Part No. 43072), provided with Model 252, to provide a true four-terminal connection to the unknown or test fixture.

3. L, R, C, and G Pushbuttons -- Function pushbuttons select the type of meter circuit that will measure series inductance (L) and resistance (R) or parallel capacitance (C) and conductance (G).

4. D Pushbutton -- The push-to-read D (dissipation factor) pushbutton displays the D of a capacitor or inductor when the C or L function is selected.

5. Range Switch -- Selects the decimal multiplier and units of measurement for the meter circuit being used. The basic multipliers and units are: H (henrys), mH (millihenrys), μH (microhenrys), μF (microfarads), nF (nanofarads), pF (picofarads), kΩ (kilohms), Ω (ohms), mS (millisiemens), and μS (microsiemens).

6. Display -- A 3-1/2 digit readout for all functions.

7. Tiltstand Handle -- Aids portability; tilts instrument for easier viewing of the LED display.

2.1.2 Rear Panel



Figure 2-2. Model 252 Rear Panel

1. Bias Terminals -- Allows application of a 2V to 50VDC, 0.1A maximum bias to the capacitor being measured. (The shorting bar must be in place when not using bias feature.) See Section 2.3.6 before using external bias.

2. Output Terminals -- Provides two analog signals proporational to the function selected (L, R, C, or G) and D (for L and C). These terminals can be used with external DVM's (for increased full scale readings or resolution capability), with chart recorders, or with limits comparators, such as the ESI Model 1412B.

3. Fuse -- A 1/16A, 250V, type MDL Slow Blow for 110V line voltage, or a 1/20A, 250V, type MDL Slow Blow for 220V line voltage.

2.2 OPERATING PROCEDURE

2.2.1 Power Requirements

Before turning the power ON, make sure the instrument is set to the proper line voltage. The Model 252 contains an internal slide switch to select the nominal line voltage (see Figure 2-3). In its up position, the switch selects 100 to 125VAC, 50/60Hz operation. In the down position it selects 200 to 250VAC, 50/60Hz operation.

CAUTION

WHEN CHANGING FROM 110VAC OPERATION TO 220VAC OPERATION (OR IN OPPOSITE ORDER), BE SURE TO REPLACE THE REAR PANEL AC FUSE WITH THE PROPER VALUE FOR THE LINE VOLTAGE SELECTED.



Figure 2-3. Line Voltage Switch

Because of differing power requirements, all instruments shipped outside the United States are without a power cord connector. When placing a connector on the power cord, care must be taken to assure the wires are connected properly. The green or green with yellow stripe wire is always connected to earth ground. The white or light blue wire is connected to the neutral side of the power line. And, the black or brown wire is connected to the high side of the power line.

CONDUCTOR	COLOR	ALTERNATE COLOR
Ungrounded Line	Black	Brown
Neutral Grounding	White	Light Blue
Earth Grounding	Green	Green-Yellow



Figure 2-4. Power Cord

2.2.2 Applying Power

The push-on, push-off, ON/OFF button in its depressed position applies power to the measurement circuitry. When power is applied, the LED display lights and reads zero when in C and G modes or the display is blank with the decimal point lit in the L and R modes.

2.2.3 Connection to Unknowns

The KELVIN KLIP® test lead set (Part No. 43072) is plugged into the 252's front panel UNKNOWN connector. The test leads connector cover is spring-locking and should be squeezed together before inserting or removing.

The test leads provide a shielded, four-terminal connection to the unknown (see Figure 2-5a). The clip with the red hinge-spool provides the HI DRIVE and HI SENSE connections to the unknown and the clip with the black hinge-spool make the LO DRIVE and LO SENSE connections.



Figure 2-5a.



Figure 2-5b.

For connection to three terminal unknowns (the third connection is to ground), a ground wire must be added to the test lead set. This wire is connected as in Figure 2-5b.

2.2.3.1 Test Fixture Compensation

The Model 252 uses a 3.3pF capacitor (C6 in Figure 2-6) to compensate for the capacitance of the test leads. If the 252 is used with a test fixture, the larger capacitance of the fixture must also be compensated for. There are two methods for compensating this larger capacitance. The first method is to make a zero capacitance measurement with the test fixture connected. This reading is mentally subtracted from all other measurements.

The second method for compensating the larger capacitance is to change the value of C6 from 3.3pF to 10pF and to add an external trim capacitor. The external trim capacitor is connected in parallel with the unknown and should be of such a value that the test fixture capacitance can be trimmed to zero. Typically, the maximum value of this trim capacitor is 15pF. The trim capacitor can either be added to the test cable, between Terminals 1 and 3 of the connector as shown in Figure 2-6, or it can be added to the test fixture.



Figure 2-6. Test Fixture Compensation

2.2.4 Function Selection

Model 252 is designed to measure series inductance (L), parallel capacitance (C), series resistance (R), parallel conductance (G), and dissipation factor (D) of inductance and capacitance. One of these measurement modes is selected by depressing the proper front panel button. (If all buttons are in their out position, the instrument reverts to the inductance measurement mode.)

Example: Select the capacitance measurement function. (Depress the C pushbutton.)



Example: Select the conductance measurement function. (Depress the G pushbutton.)



If a negative sign appears on the display when measuring an unknown, possible causes are:

-- "C" button has been pushed when measuring an inductor.

-- "L" button has been pushed when measuring a capacitor.

-- The unknown is more capacitive (or inductive) than suspected. For example, an inductor that resonates below 1kHz will measure as though it were a capacitor.

-- A diode requires bias voltage (a negative reading may appear if bias voltage is not applied).

To measure the dissipation factor (D) of a capacitor or inductor, push and hold the D button while the C or L function is selected. (The display will blank if D is pushed, while the unit is in R or G mode and the unknown is resistive.)

When measuring the conductance (G) of a capacitor or the resistance (R) of an inductor, the R or G reading may be quite small. In trying to acquire more resolution by moving the range switch to the next lower range, the display will usually blank, indicating overload. Blanking occurs because the instrument is still monitoring the reactive component (L or C) in addition to measuring and displaying the loss component (R or G). In moving to the next lower range, the reactive component becomes too large for the lower range causing overload and a blanked display.

Example: While measuring a Ø.82µF capacitor using the 2µF range, the 252 display reads:



Switching to the conductance (G) mode, the display may read:



Trying further to investigate this reading, you down range from the 20mS range to the 2000µS range. The display blanks:



The display blanks because, while ranging from 20mS to 2000µS, the capacitance range changed from 2µF to 200nF. Since 0.82µF is too large to measure on the 200nF range, an overload occurs and the display blanks.

2.2.5 Range Selection

Measurement ranges are selected by a seven-position rotary switch. See Figure 2-7 for full-scale range values. For maximum display resolution and accuracy, select the range that gives the largest on-scale reading. The position of the range switch indicates the readout units (i.e. µH, mH, H, etc.). If either the reactive or resistive value of the unknown exceeds the range selected, the display will blank.

		Ranges						
		0	1	2	3	4	5	6
Full Scale Values	L_S	200 µH	2 mH	20 mH	200 mH	2 H	20 H	200 H
	C_P	200 pF	2 nF	20 nF	200 nF	2 µF	20 µF	200 µF
	R_S	2 Ω	20 Ω	200 Ω	2 kΩ	20 kΩ	200 kΩ	2000 kΩ
	G_P	2 µS	20 µS	200 µS	2 mS	20 mS	200 mS	2000 mS
	D	1.999						

Figure 2-7. Range Chart

Example: Set the instrument to measure 100pF capacitors.



Example: Set the instrument to measure 100kΩ resistors.



If the value of a component is not known, set the instrument to its highest range before connecting the test leads. Step the instrument down one range at a time until the range with the highest resolution is reached. If the display is blank after the component is connected, the value of the component is too large to measure directly, see Section 2.3 for alternate measurement techniques.

2.2.6 Summary of Operation

Press the appropriate function button (L, C, R, or G) and connect the KELVIN KLIPS® to the unknown. For maximum accuracy, select the range that gives the largest on-scale reading. The measurement is displayed after one second or less. Repetitive measurements are made at the rate of four per second. The position of the range switch indicates the readout units (i.e. μH, mH, H, etc.). If the value of the unknown

exceeds the range selected, the display will remain blanked. To measure the dissipation factor of a capacitor or inductor, push and hold the D button while the C or L function is selected.

2.3 MEASUREMENT TECHNIQUES

2.3.1 Resistance Measurements at 1kHz

When using the resistance function, the maximum measurement is 1999kΩ. For measurements from 1999kΩ to 1000MΩ, use the G (conductance) function where G=1/R or R=1/G. Measurement accuracy is determined by ±(0.25% + 1 digit) where 1 digit will be ±100% at 1000MΩ, ±10% at 100MΩ, and ±1% at 10MΩ.

High-value, bobbin-type, wire-wound resistors have shunt capacitance which may affect measurements by appearing as "$-L_s$". This causes some difference between AC and DC resistance measurement values.

The resistive component (Rs) of inductors and transformers, when measured at 1kHz, will differ from the DC winding resistance. If the core material is air or low loss (e.g. powdered permalloy), the AC and DC values will be nearly equal. For laminated magnetic cores, the AC resistance will be greater than the winding resistance (sometimes much greater). This is caused by the core losses which change both with frequency and test current level.

2.3.2 Capacitance Measurements

Shielded devices. For shielded devices, the shield should be connected to the rear panel -Bias terminal to provide proper guarding.

Most two- and four-terminal cables or sorting fixtures are not guarded immediately adjacent to where the unknown capacitor is connected, and a zero capacitance reading can change by a few picofarads whenever the ends of the cable or fixture terminals are physically moved. Always maintain the physical orientation of the leads or terminals after checking for zero capacitance. Subtract zero-capacitance reading from the capacitance reading for the unknown.

Equivalent Series Resistance (ESR). To measure series resistance of capacitors, use the Rs function. Series resistance will be displayed on the front panel.

Series Capacitance. The Ls function can be used to measure series capacitance. (The capacitor is measured with a constant current rather than a constant voltage.) A negative reading will be shown, and the following equation must be used to convert the "-Ls" readings to the "Cs" values:

$$Cs = \frac{1}{(2\pi f)^2 (Ls)}$$

Where: f = frequency in hertz
Cs = series capacitance in farads
Ls = series inductance in henrys (ignore sign)

Dissipation factor (D). The D of large value electrolytic capacitors can be determined using the Ls and Rs functions using the same range. The following equation converts the Rs and Ls, in counts (ignore the sign and decimal point), to D values:

$$D = 1.59 \times \left(\frac{Rs}{Ls}\right) \text{ for 100Hz and 1kHz operation}$$

$$D = 1.33 \times \left(\frac{Rs}{Ls}\right) \text{ for 120Hz operation}$$

Where: D = dissipation factor
Rs = equivalent series resistance reading in counts
Ls = series inductance in counts

Parallel Capacitance Measurements. The relation between parallel and series capacitance measurements is:

$$Cs = (1 + D^2)\ Cp \quad \text{where } D = Gp/2\pi fCp$$

The following graph illustrates the relation between Cs and Cp. The graph is used to obtain a Cs value for a capacitor by measuring the Cp and D values of the capacitor. To locate the Cs value, find the measured D value along the horizontal axis of the graph. (The D reading from the Model 252 will be a Dissipation Factor value not a Percent Dissipation value.) Move up the graph to the B curve and read the corresponding value on the righthand vertical axis. Multiply the measured Cp value by this reading to obtain the desired Cs value.

The A curve on the graph is used with the vertical axis in the same manner as described above. The A curve and lefthand scale provide the Percent Difference between Cs and Cp.



2.3.3 Inductance Measurements

The 252 measures the total impedance connected to its terminals. Both the unknown inductor and its leads contribute to this impedance. The leads have some resistance and inductance which affect the value read from the meter.

In making high inductance measurements, avoid AC pickup and keep the stray capacitance to a minimum. To minimize both effects, keep hands as far as possible from the inductor being measured. Keep the leads as short and direct as possible. Take care to avoid coupling stray magnetic fields into the inductor.

For greatest accuracy on low inductance measurements, minimize the lead impedance. Closely spaced, twisted leads will reduce the inductance and pickup of stray fields. The KELVIN KLIPS® Cable Assembly will add about 0.5μH to Ls and 0mΩ to Rs. Short the test leads together and subtract the reading from the unknown inductance measured.

Measuring leakage inductance of transformer windings is an easy task for the Model 252 since there are no "false nulls" as found in manually-balanced bridges.

2.3.4 Determining Quality Factor (Q) of Inductors

To determine the Q of inductors, push the D button, record the D reading and calculate Q from the equation Q=1/D. This procedure works well for Q≥Ø.5 (or D≤2). For Q≤Ø.5, measure the inductance of the unknown and record the readout display in digits only. With the RANGE switch in the same position, press the R button and note the readout display in digits, only. Do not record the decimal point or range multiplier for Ls or Rs; they are not required. Turn to Figure 2-8 for the D-Q nomograph. Use a straightedge to line up the Ls digits recorded (left-hand scale) with the Rs digits recorded (right-hand scale). Q is taken from the left-hand side of the center scale where the straightedge crosses. For example, with an Ls digital display of "3ØØ" and an Rs digital display of "6ØØ", Q is Ø.3. Accuracy of Q is primarily limited by interpolation of the nomograph.

Although the nomograph provides a convenient way to approximate Q, it may be determined more accurately by the formula:

$$Q = Ø.628 \times \left(\frac{Ls}{Rs}\right) \text{ for 1ØØHz and 1kHz operation}$$

$$Q = Ø.754 \times \left(\frac{Ls}{Rs}\right) \text{ for 12ØHz operation}$$

Where: Q = Quality factor
Rs = equivalent series resistance reading in counts
Ls = series inductance in counts

See Table 1-1 for accuracy of Ls and Rs measurements.

2.3.5 Determining Dissipation Factor (D) of Capacitors Using Nomograph

Model 252 will measure dissipation factor up to 1.999. For D values greater than 2 the nomograph should be used. Measure the capacitance of the unknown and record the readout display in digits only; do not record the decimal point. With the RANGE switch left in the same position, push the G button and note the readout display in digits only; do not record the decimal point. Turn to Figure 2-8 for the D-Q nomograph. Use a straightedge to line up the Cp digits recorded (left-hand scale) with the Gp digits recorded (right-hand scale). D is taken from the right-hand side of the center scale where the straightedge crosses. For example, with a Cp digital display of "200" and a Gp digital display of "300", D is 2.3. Accuracy of D is primarily limited by interpolation of the nomograph.

Dissipation factor may be determined more accurately by the formula:

$$D = 1.59 \times \left(\frac{Gp}{Cp}\right) \text{ for 100Hz and 1kHz operation}$$

$$D = 1.33 \times \left(\frac{Gp}{Cp}\right) \text{ for 120Hz operation}$$

Where: D = dissipation factor
Gp = parallel conductance reading in counts
Cp = parallel capacitance in counts

See Table 1-1 for accuracy of Cp and Gp measurements.

Electrolytic capacitors with high dissipation are usually unstable and the measured values may drift.



Figure 2-8. Model 252 D-Q Nomograph

2.3.6 Using the Bias Feature

WARNING

TO AVOID PERSONAL INJURY FROM ELECTRICAL SHOCK OBSERVE SAFETY PROCEDURES AS NEEDED FOR THE BIAS VOLTAGE APPLIED.

Bias voltage is necessary or desirable when measuring the capacitance of diodes or of some electrolytic or tantalum capacitors.

There is a provision on the rear panel for connecting an external bias source of zero to 50VDC. Remove the shorting link between BIAS terminals when using this feature.

If the external bias supply is not a low AC impedance type, it should have a capacitor bypass of five times the range full-scale value connected across the bias terminals. The bias supply should be current limited to 0.1A maximum.

With no bias supply connected, the rear-panel + and - terminals must be strapped together.



Figure 2-9. Rear Panel

2.3.7 Signal Output Terminals

Terminals are provided on the Model 252's rear panel for connecting external devices: digital voltmeters to increase full-scale resolution or simultaneous D reading; chart recorders to evaluate unknowns with changing impedances; limits comparators (ESI Model 1412B) for HI, LO, GO, and HI D component sorting.

The analog output signals are proportional to the function selected (L, C, R, or G) and D (for L and C only). Specifications for these signals are:

Scale Factor	1V per 1000 counts
Source Resistance	1000Ω
Accuracy	±(0.25% of reading + 1mV)

2.3.8 Measuring Grounded Unknowns

The Model 252 is designed to measure ungrounded components. However, with a minor modification, it can be used to measure impedances that are connected directly or tightly coupled to earth ground.

This first procedure outlines the modification of all instruments with serial numbers 928999 and smaller.

1. Remove all instrument power.

2. Remove instrument cover by unscrewing the two cover screws and sliding the cover off.

WARNING

TO AVOID PERSONAL INJURY FROM ELECTRIC SHOCK DO NOT REMOVE INSTRUMENT COVERS OR PERFORM ANY MAINTENANCE OTHER THAN DESCRIBED IN THIS MANUAL. INSTALLATION AND MAINTENANCE PROCEDURES DESCRIBED IN THIS MANUAL ARE TO BE PERFORMED BY QUALIFIED SERVICE PERSONNEL ONLY.

3. Locate the Power Supply Circuit Assembly, Part No. 53157. (Attached to the rear panel.)

4. Remove jumper wire W1 from this assembly (see Figure 2-10).



Figure 2-10. Power Supply Circuit Assembly

5. Locate the Range Switch on the front panel.

6. Unsolder and isolate the ground solder lug from the Range Switch terminal and black wire, making sure that the black wire remains attached to the Range Switch terminal (see Figure 2-11).



Figure 2-11.

7. Replace instrument cover.

The modification of instruments with Serial Numbers 929000 and greater is as follows:

1. Locate the connector strip on the rear panel (see Figure 2-12).

2. Remove and isolate the spade lug from the terminal marked COM.



Figure 2-12.

To make measurements of grounded unknowns, connect the grounded side of the unknown to the KELVIN KLIPS• with the red insulating hinge-spool and the ungrounded end to the clip with the black spool (see Figure 2-13).



Figure 2-13.

A small zero offset may become apparent during low value grounded capacitance measurements. This offset can be corrected by making a reading with the unknown connected only to the red clips and subtracting it from the final measured capacitance.

To return to normal ungrounded measurements reverse the above procedures.

2.3.9 Measuring Battery Impedance

Current producing capabilities of a battery can be determined by using the 252 to measure its impedance. Series resistance and series capacitance measurements can be made on batteries with voltages up to 50V. To measure series resistance:

1. Set the 252 to measure Rs and to the highest measurement range.

2. Remove the shorting bar from the + and - bias terminals. The bias terminals are located on the instrument's rear panel terminal strip.

3. Connect a 500µF (or larger) capacitor, whose working voltage is greater than the rated voltage of the battery, to the bias terminals (see Figure 2-14). Observe correct polarity when connecting the capacitor.

4. Connect the test leads with the red hinge-spool to the positive battery terminal and the lead with the black hinge-spool to the negative terminal (see Figure 2-14).

5. Move the range switch to the range with the highest resolution displayed (see Section 2.2.5) unless that range uses excessive test current for the size of battery being measured. See Table 1-1 for range test currents.



Figure 2-14. Measuring Battery Impedance

The Ls function must be used to measure the series capacitance of a battery. Instrument set up is the same as for Rs measurements (Figure 2-14). A negative reading will be shown, and the following equation must be used to convert the "-Ls" reading to "+Cs" values:

$$Cs = \frac{1}{(2\pi f)^2 \ (-Ls)}$$

Where: f = frequency in hertz
Cs = series capacitance in farads
Ls = series inductance in henrys

Care must be taken so the measurement current level of the 252's internal source does not exceed the rating of the battery being tested. See Table 1-1 for range test currents.

2.3.10 Component Sorting

Model 252, when used with Model 1412B Limits Comparator, sorts components according to percent deviation and dissipation factor, simultaneously. Component sorting can be manual or automatic; Model 1412B provides four contact closures for operating automatic component handling equipment.

Model 1412B front panel controls set upper and lower limits around a nominal value (L, R, C, or G), with either of two full scale ranges -- 10% or 100%. An upper limit for dissipation factor (D) is also set by a front panel control. Test results are shown by four front panel LEDs. They indicate:

GO indicates value between - and + % limits and less than D limit.

LO indicates value less than - % limit.

HI indicates value greater than + % limit.

D HI indicates dissipation factor greater than limit.

Connect the 252 and 1412B as shown in Figure 2-15.




pin number J1
L,R,C,G INPUT 1 19 +15V
D INPUT 2 20 -15V
V REF 3 21 COM
COM 4 22 COM
COM 5 23
% OUTPUT 6 24
7 25
LO 8 26 LO
LO 9 27 LO
10 28
HI 11 29 HI
HI 12 30 HI
13 31
GO 14 32 GO
GO 15 33 GO
16 34
D HI 17 35 D HI
D HI 18 36 D HI
LO 8 — 26 LO
LO 9 — 27 LO
HI 11 — 29 HI
HI 12 — 30 HI
GO 14 — 32 GO
GO 15 — 33 GO
D HI 17 — 35 D HI
D HI 18 — 36 D HI

NOTE: For resistive unknowns connect the D lead to the COM terminal

Figure 2-15 Model 252 and Model 1412B Interconnection

Operation

The following instructions assume Model 252 and Model 1412B are connected and both are turned on.

Model 252 setup: (for capacitor sorting)

1. Push the C button on the Model 252 front panel.

2. Set the range switch to give maximum resolution on the display.

Model 1412B setup:

1. Set the rear panel V REF switch to INTERNAL.

2. Set NOMINAL VALUE switches to anticipated value.

3. Set LIMITS FULL SCALE toggle switches, and % LIMITS and D LIMIT dials to desired limits.

The measurement:

Connect unknown to the Model 252. If the unknown is within the set limits, the GO LED will light. If it is not within limits, the appropriate indicator will light.

Example: Sort 15µF capacitors within +25%, -15% limits with a maximum dissipation limit of 0.0250. Instrument set up is as follows:

Model 252:

Function and Range settings



Model 1412B:

Nominal value setting



- % Limits setting

+ % Limits setting

LIMITS FULL SCALE Toggle switches settings

D Limit setting

Connect the unknown and observe indicator lights. If the LO LED is lit, the unknown is lower than the lower % limit. Similarly, if the HI LED is lit, the unknown is a larger value than the upper % limit. If the D HI LED is lit, the dissipation factor of the unknown exceeds the set limit. If the GO indicator is lit, the unknown value is within the desired limits; in the example, 15μF, +25%, -15%, with a dissipation factor less than Ø.Ø25Ø.

For further operating instructions, see the Model 1412B instruction manual.

SECTION 3
CIRCUIT DESCRIPTIONS

3.1 GENERAL DESCRIPTION

When measuring Cp or Gp, a voltage is applied across the unknown and range resistor in Figure 3-1. The voltage across the unknown is held constant to within one part in several thousand by the 1kHz oscillator feedback control circuit. A current proportional to the value of the unknown impedance is produced through the range resistor. The resultant voltage drop across the range resistor is separated into two vector components by the Phase Sensitive Detectors (PSD) and Reference Voltage Generator (RVG). Receiving gating signals from RVG, the PSD will measure Gp, the component in phase (0°) with RVG, and Cp, the component at quadrature (90°). The selected detector output is then fed to the A/D Converter which has a readout of 1000 counts/volt.

When measuring Ls or Rs, a voltage is applied across the unknown and range resistor in Figure 3-1. The voltage across the range resistor is held constant to within one part in several thousand by the 1kHz oscillator feedback control circuit. With a constant current flowing through the unknown, a voltage proportional to the value of the unknown impedance is developed across the unknown. The resultant voltage drop is separated into two vector components by the Phase Sensitive Detector (PSD) and Reference Voltage Generator (RVG). Receiving gating signals from the RVG, the PSD will measure Rs, the component in phase (0°) with the RVG, and Ls, the component at quadrature (90°). The selected detector output is then fed to the A/D Converter which has a readout of 1000 counts/volt.



Figure 3-1. Model 252 Simplified Diagram

When measuring the dissipation factor (D) of inductors or capacitors, an analog ratio circuit is used. This circuit compares the loss component (Rs or Gp) to the reactive component (Ls or Cp) and presents a voltage proportional to D to the A/D Converter.

Analog output voltages are available at the rear of Model 252 for the selected function (L, C, R or G) and simultaneous D (for L or C unknowns).

3.2 DETAILED CIRCUIT DESCRIPTIONS

In the following descriptions several conventional resistor networks are involved. To identify which resistor elements are being discussed, the elements are identified by the lower terminal numbers and each is prefixed by a small letter "r". Example: U7 (r1 and r2).

3.2.1 1kHz Oscillator

The 1kHz signal applied to the unknown and range resistor is generated by a power op amp U2, used as a Wien bridge oscillator, with C20, C21 and U19 (r4 and r5) controlling the frequency. Amplitude control for the oscillator is provided by U27 working as a reference voltage null integrator and U35 as the feedback control element. Transformer T2 provides coupling to the unknown and range resistor with provision for inserting bias voltage at rear terminals.

3.2.2 Unknown Differential Amplifier

The voltage drop across the unknown is measured by differential amplifier U3-3. Amplifiers U1-3 and U1-4 act as high impedance AC-coupled buffers to prevent loading the unknown. Resistors U6 (r7 and r8) and U7 (r1 and r2) determine the gain (x1) of U3-3. Trimmer R6 is an adjustment for common-mode voltage rejection. Amplifiers U3-2 and U3-3 are switch (U5) selectable gain stages that provide an overall gain of x1, x10 or x100. Trimmers R11 through R17, switched by U4, provide phase adjustments for each range.

3.2.3 Range Amplifier

The voltage drop across the range resistor (R22, R24 through R27), selected by U12, is measured by amplifier U1-2. Relay K1 switches R22 in series with the unknown for ranges 20Ω and 2Ω (or equivalent). For these ranges U36 is used as a gain amplifier of x1 or x10. Other ranges use U36 as a guard amplifier to reduce the effect of shunt impedances on high value unknowns.

3.2.4 Reference Generator and Phase Sensitive Detector

Reference Generator

A DC voltage proportional to the voltage across the unknown or range resistor is generated by U13-2 and U13-3 used in an operational rectifier circuit. For reactive unknowns, amplifier U13-1 is used as the AC voltage reference input to the operational rectifier. Gating signals for the in-phase and quadrature phase sensitive detectors are produced by U15-3 and U15-4 used as zero crossing detectors. Trimmer R35 is used for adjustment of the 90° reference gate.

Phase Sensitive Detectors

Voltage across the unknown (Rs or Ls) or across the range resistor (Cp or Gp) is measured by two phase-sensitive detectors. The in-phase (0°) detector consists of amplifiers U17-1 and U30, together with resistor network U23 and switching gate U20. Ripple filtering is provided by R46, R47, C17, and C18. The quadrature (90°) detector consists of amplifiers U17-4 and U22, together with resistor network U21 and switching gate U20. Ripple filtering is provided by R41, R42, C13, C14, and C15. Trimmer R45 is used to adjust the full scale value of L or C.

3.2.5 Dissipation Factor Circuit

The dissipation factor (D) of inductors or capacitors is derived from the ratio of the loss component (0° phase detector) and reactive component (90° phase detector). This is done with a pulse-width modulator consisting of integrating amplifier U26 and resistor summing network using U28, R60 and R58. The ratio of R (or G) to L (or C) changes the pulse width ratio of the switching amplifier U27 (with switch U24) to produce a null at the integrator input. Comparator U15-2 acts as the switch driver and is frequency set by the test frequency of the instrument. Amplifiers U27-1 and U27-3 together with

switch U24 are the demodulator which produces an analog voltage proportional to the pulse width variation and scaled to equal the dissipation factor. Trimmer R52 adjusts full scale D value.

3.2.6 Overload Blanking Circuit

U15-1 is a comparator amplifier. It monitors the signal levels of the phase detector input and the level regulator amplifier U27-4. When either level exceeds its limit, comparator U15 goes low. This output is combined with a digital overrange signal at the input of a NAND gate of U29 to provide a display blanking output.

3.2.7 Analog to Digital Converter

The A/D converter resides in U31 (MC14433) which is a CMOS dual-slope integrating converter. This circuit provides a multiplexed output which is decoded and displayed by a plug-in circuit assembly (42916).

3.2.8 Function and Range Programming

A bipolar PROM (U33) stores the necessary range and function programming information for Model 252. The stored program has 32 addresses each with 8 bits of data. These addresses are reached by 2 function bits and 3 range bits which are set by the front panel switches.

SECTION 4
MAINTENANCE

WARNING

TO AVOID PERSONAL INJURY FROM ELECTRIC SHOCK DO NOT REMOVE INSTRUMENT COVERS OR PERFORM ANY MAINTENANCE OTHER THAN DESCRIBED IN THIS MANUAL. INSTALLATION AND MAINTENANCE PROCEDURES DESCRIBED IN THIS MANUAL ARE TO BE PERFORMED BY QUALIFIED SERVICE PERSONNEL ONLY.

4.1 CALIBRATION PROCEDURE

Although Model 252 is inherently stable, good maintenance practices suggest it be calibrated every six months; more often if used in extreme environments or if inaccurate readings are suspected.

Equipment Required	Minimum Specifications
Frequency Counter	Measure 1000Hz with resolution of 1Hz
Digital Voltmeter	Resolution of 100µVDC
Resistance Standard	1000Ω ±0.05%, non-inductive such as ESI Model SR 1.
Capacitance Standard	100nF, ±0.05%, silvered mica or polystyrene
Dissipation Factor Standards	1µF Decade Capacitor, 5 dial, polystyrene, D<0.0002; 10µF polystyrene capacitor with D known to ±0.001; 100µF capacitor with D known to ±0.001
KELVIN KLIPS® Four-Terminal Clips	ESI Part No. 43072

Refer to Figure 4-1 for adjustment locations.

4.1.1 Power Supply Check

Setup - Turn on instrument and allow ten minute warm-up period; then remove cover. Check - Power supply voltages at C44 and C45 should be + and -5V +/-5% (4.75V to 5.25V). Use rear panel COM terminal for voltmeter low input.

4.1.2 Oscillator Adjust

Setup - Press R function button. Set RANGE switch to 2000 ohms and short test leads together. Connect voltmeter high input to Test Socket (TS) pin (2) and frequency counter input to TS (1). Use rear panel COM terminal for circuit common. Adjustment - Set trimmer S for DVM reading of 0V (+/-0.2V). Counter reading should be 1000Hz (+/-10Hz).

4.1.3 L, C, R, G Alignment

1. R, G PHASE DETECTOR ZERO TRIM. Setup - Press R function button and set RANGE switch to 2000 ohms and short test leads together. Connect DVM input to TS (6). Adjustment - Set trimmer P for +0.3mV +/-0.1mV on DVM.

2. UNKNOWN AMP COMMON MODE TRIM. Setup - Press R function button and set RANGE switch to 2 ohms and short test leads together. Adjustment - Adjust trimmer J for DVM reading of 0V +/-0.3mV at TS (6).

3. L, C PHASE DETECTOR ZERO TRIM. Setup - Press L function button and set RANGE switch to 200mH. Short test leads together and connect DVM input to TS (5). Adjustment - Set trimmer M for +0.3mV +/-0.1mV on DVM.

4. L, C, R, G OUTPUT AMP ZERO TRIM. Setup - Same as step 3 except connect DVM input to TS (4). Adjustment - Set trimmer V for ØV ±Ø.3mV on DVM.

5. UNKNOWN AND RANGE AMP PHASE TRIM. Setup - Press C function button and set RANGE switch to 200nF. Connect decade capacitor to unknown test leads and set at 160nF (±5%). Adjustment - Note voltage at TS (6) and then push L function button, again noting TS (6) voltage. The two should be equal (±Ø.3mV). If not, adjust trimmer X until they are. Next adjust trimmer D for ØV ±Ø.3mV output at TS (6) for both L and C functions.

6. QUADRATURE GATE PHASE TRIM. Setup - Press R function button and set RANGE switch to 2ØØØΩ. Connect 1ØØØΩ (±Ø.Ø5%) standard to unknown test leads. Adjustment - Set trimmer L for DVM reading of ØV ±Ø.3mV at TS (5).

7. R, G FULL SCALE TRIM. Setup - Same as step 6. Adjustment - Set trimmer K for front panel digital reading of 1ØØØ. With DVM, read voltage at TS (4) and adjust trimmer R for +1.ØV (±Ø.5mV).

8. L, C FULL SCALE TRIM. Setup - Press C function button and set RANGE switch to 200nF. Connect 100nF (±Ø.Ø5%) standard to the unknown test leads. Adjustment - Set trimmer N for front panel reading of 1ØØØ.

4.1.4 Dissipation Factor (D) Alignment

1. LOW LEVEL ZERO TRIM. Setup - Press C function button and set RANGE switch to 200nF. Connect a low loss (D<Ø.ØØØ2) decade capacitor to unknown test leads. Adjustment - Set the decade capacitor to 105nF and, with a DVM, note the voltage at TS (3) or D output terminal at the rear panel. Set the decade to 5nF and

again note the voltage. If the two voltages are not equal (+/-1mV), adjust trimmer W until switching between 105nF and 5nF produces equal readings.

2. HIGH LEVEL ZERO TRIM. Setup - Same as step 1. Adjustment - Set decade capacitor to 100nF and adjust trimmer U until voltage at TS (3) or D output is 0V +/-0.3mV.

3. D FULL SCALE TRIM. Setup - Same as step 1. Adjustment - Set the decade capacitor for a front panel C reading of 159.2nF. Connect the 1000 ohm standard resistor in parallel with the decade capacitor. Set trimmer T for a front panel D reading (push D button) of 1.000.

4. 200pF RANGE D TRIM. Setup - Press C function button and set RANGE switch to 200pF. Connect the low loss (D<0.0002) decade capacitor to unknown test leads. Adjustment - Set the decade to 100pF (+/-20%) and adjust trimmer H for 0V +/-0.3mV at TS (3), or D output.

5. 2nF RANGE D TRIM. Setup - Same as step 4, except set RANGE switch to 2nF. Adjustment - Set decade to 1nF (+/-20%) and adjust trimmer F for 0V +/-0.3mV at TS (3), or D output.

6. 20nF RANGE D TRIM. Setup - Same as step 4, except set RANGE switch to 20nF. Adjustment - Set decade to 10nF (+/-20%) and adjust trimmer E for 0V +/-0.3mV at TS (3), or D output.

7. 2uF RANGE D TRIM. Setup - Same as step 4, except set RANGE switch to 2uF. Adjustment - Set decade to 1uF (+/-20%) and adjust trimmer C for 0V +/-0.3mV at TS (3), or D output.

8. 20uF RANGE D TRIM. Setup - Set RANGE switch to 20uF and connect a 10uF (+/-20%) capacitor (D known to +/-0.001) to unknown test leads.

   Adjustment - Adjust trimmer B for front panel D reading (push D button) equal to D value of the 10uF capacitor.

9. 200uH RANGE D TRIM. Setup - Press the L function switch and set the RANGE switch to 200uH. Short the Kelvin Klips together and note the zero inductance (approximately +5 counts) for use in the adjustment. Connect a 1.2 ohm, 5% carbon composition resistor to the test leads.

   Adjustment - Adjust trimmer A for a display equal to, or -1 count from the zero inductance value noted when the Kelvin Klips were shorted together.


D G R C L
S6 S5 S4 S3 S2
J1
R2 R1
CR16 R8 R3
C4 C1
CR17 R18
K1
U10
CR15 C25 C26
U31
C24
R28 Q1 CR8
R9 CR5 R4
C28 R78 R77 C27
R22 R23 R29
CR4 CR2 CR1
R24 R26 R25
C5 R10 C2 R5
U34 U33 U32
J5 J4 U29
R82
U13 C37 U9
U12
R27 C35
CR6 CR7
CR10 CR9
C6 U36 U1
R79 R83 C8 C7
R48 R49
U23 U21
U14 U4
A R11
R19 R20 R21 CR3
B H
C3
C30 C10 C11 R32 R33 R30
C F
U5
D
U11 U6
C46
E C33 J R6
R7
K
C9 R76 R75 R39 R38 R37 R36
U17 C16 U20 C39 R84
X
U
U16 U3 U8 C31 U7
C12 R40 C29 CR12 R73
U15
R34 R74 R72
R66 C41
C38 C34
C19 C36
R65 R
U24 R53 U19
R67 R68 R69 R70
L
R63 R64 R59 CR11
R44 R80 R41 U18 C23
R44 C14 C15 C13 R42
U22 U26
R51
T2
R58 U28 C22 U27 C21 C20 U35 C43 C42
R61
N
U30 U25
T
CR18
M
C17 R60 R47 C18
R87
S
P CR14 R46
W
CR13 R86
V C40
U2
R56 R55
C44 C45

Figure 4-1.

4.2 MAINTENANCE

This section of the manual contains maintenance information for use in preventive maintenance, corrective maintenance, and troubleshooting of the Model 252.

4.2.1 Preventive Maintenance

WARNING

REMOVAL OF INSTRUMENT COVERS MAY CONSTITUTE AN ELECTRICAL HAZARD AND SHOULD BE ACCOMPLISHED BY QUALIFIED SERVICE PERSONNEL ONLY.

Preventive maintenance performed on a regular basis will improve the reliability of this instrument. It may include cleaning, visual inspection, or even monitoring the operating environment.

4.2.1.1 Cleaning

CAUTION

AVOID THE USE OF CHEMICAL CLEANING AGENTS WHICH MIGHT DAMAGE THE PLASTICS USED IN THIS UNIT. DO NOT APPLY ANY SOLVENT CONTAINING KETONES, ESTERS OR HALOGENATED HYDROCARBONS. TO CLEAN, USE ONLY WATER SOLUBLE DETERGENTS, ETHYL, METHYL, OR ISOPROPYL ALCOHOL.

Exterior. Loose dust may be removed with a soft cloth or a dry brush. Water and mild detergent may be used; however, abrasive cleaners should not be used.

Interior. Use low-velocity compressed air to blow off the accumulated dust. Hardened dirt can be removed with a cotton-tipped swab, soft, dry cloth, or a cloth dampened with a mild detergent and water.

4.2.1.2 Visual Inspection

This instrument should be inspected occasionally for such defects as broken connections, improperly seated semiconductors, damaged circuit boards, and heat-damaged parts.

The corrective procedure for most visible defects is obvious. If heat damaged components are found, particular care must be taken. Overheating usually indicates other trouble may be present in the instrument. It is important that the cause of overheating be corrected to prevent recurrence of the damage.

4.2.2 Troubleshooting

The following troubleshooting information is provided to augment other sections of this manual. The Circuit Description and Part Lists and Schematic Diagrams sections should be used to full advantage. Section 3 in this manual gives circuit description information while, Section 5 contains the part lists and schematic diagrams.

4.2.2.1 Troubleshooting Aids

Schematic Diagrams. Schematic diagrams are provided on foldout pages in Section 5. The electrical value and circuit numbers of each component are shown on the diagrams. Power supply voltages are also shown.

Circuit-Board Illustrations. Illustrations of circuit boards are shown along with the schematic diagrams. Each board-mounted electrical component is identified by its circuit number.

Test Point Locations. Test point locations have been indicated on both the schematic diagrams and the circuit-board illustrations.

Component Color Code. Colored stripes or dots on resistors and capacitors signify electrical values, tolerances, etc., according to the EIA standard color code. Components not color-coded usually have the value printed on the body.

Multi-pin Connector Identification. Multi-pin connectors are soldered to the circuit boards. They mate with ribbon type cable assemblies to carry signals between boards. Connector pin 1 is indexed with a number 1 etched on the circuit board. Each connector is identified by a P number and can be located by using the circuit board illustration in Section 5 of this manual. P numbers shown on the illustration correspond to the P numbers on the schematic diagrams.

4.2.2.2 Troubleshooting Procedure

This troubleshooting procedure checks the simple trouble sources before proceeding with more extensive troubleshooting. The first few checks ensure proper connection and operation. If the trouble is not located by these checks, the remaining steps aid in locating the component. When the defective component is located, it should be replaced using the information given under Corrective Maintenance.

1. **Check Instrument Set Up.** Make sure the instrument is properly plugged into a wall socket. Also, check the rear panel line voltage switch and the line fuse to see that they match the line voltage being used.

2. **Visual Check.** Visually check the portion of the instrument in which the trouble is suspected. Many problems can be located by visual indications such as unsoldered connections, broken wires, damaged circuit boards, damaged components, or components bent over and touching.

3. **Check Voltages.** A circuit stage may not be operating due to incorrect supply voltages. Typical supply voltages are given on

the diagrams; however, these are not absolute and may vary slightly between instruments.

4. **Trace the Signal.** The analog portion of the circuitry can be checked by tracing the signal with an oscilloscope. By noting where the signal disappears or distorts, the source of trouble can be located.

5. **Check Individual Components.** The following methods are provided for checking the individual components. Components which are soldered in place can sometimes be checked by disconnecting one end to isolate the measurement from the effects of surrounding circuitry.

   a. TRANSISTORS. It is always best to check transistor operation under operating conditions. Transistors that are soldered to the circuit board should first be checked in-circuit using a dynamic transistor testor; then a replacement can be substituted to verify that the old transistor is bad. Socketed transistors can be checked by substituting a component known to be good; however, be sure that circuit conditions are not such that a replacement might also be damaged. If substitute transistors are not available, check the old transistor out-of-circuit using a dynamic tester. Be sure the power is off before attempting to remove or replace any transistor.

   b. INTEGRATED CIRCUITS. Analog IC's such as comparators and operational amplifiers can usually be checked in-circuit with a voltmeter or test oscilloscope. An understanding of the device and circuit operation is essential for this type of troubleshooting. (For example, an op amp can be tested by measuring the input and output circuit voltages and comparing this ratio to the ratio of input and feedback resistors.)

      Analog IC's that are socketed can also be checked out-of-circuit using a dynamic tester such as the Model 1234.

Digital IC's are best checked in-circuit using a logic probe or voltmeter. Use care when checking voltages and waveforms around DIP (Dual-Inline-Package) IC's so that adjacent leads are not shorted together. A convenient means of connecting a test probe to 14 and 16 pin IC's is with an IC test clip. This device also doubles as an extraction tool.

c. **DIODES.** A diode can be checked for an open or shorted condition by measuring the resistance between terminals with an ohmmeter set to the R X 1k scale. The diode resistance should be very high in one direction and very low when the meter leads are reversed.

DO NOT USE AN OHMMETER SCALE THAT HAS A HIGH INTERNAL CURRENT. HIGH CURRENTS MAY DAMAGE THE DIODES UNDER TEST.

d. **RESISTORS.** Check resistors with an ohmmeter. Resistor tolerance is given in the Parts List. Resistors normally do not need to be replaced unless the measured value varies widely from the specified value.

e. **CAPACITORS.** A leaky or shorted capacitor can be detected by checking resistance with an ohmmeter on the highest scale. Use an ohmmeter that will not exceed the voltage rating of the capacitor. (Be careful to observe correct polarity when checking electrolytic capacitors.) The resistance reading should be high after initial charge of the capacitor. An open capacitor can best be detected with a capacitance meter, or by checking whether the capacitance passes AC signals.

4.3 CORRECTIVE MAINTENANCE

Corrective maintenance consists of component replacement and instrument repair.

4.3.1 Obtaining Replacement Parts

Standard Parts. All electrical and mechanical replacement parts for the Model 252 can be obtained from Tegam Incorporated. However, many of the electronic components can be obtained locally in less time than is required to order them from Tegam. Before purchasing or ordering replacement parts, check the parts list for value, tolerance rating, and description.

> NOTE: When selecting replacement parts, it is important to remember that the physical size and shape of a component may affect the performance of the instrument. All replacement parts should be direct replacements unless it is known that a different component will not adversely affect instrument performance.

Order all special parts directly from Tegam Incorporated.

4.3.2 Component Replacement

WARNING

DISCONNECT ALL POWER TO THE INSTRUMENT BEFORE REPLACING COMPONENTS. FAILURE TO DO SO MAY RESULT IN ELECTRICAL SHOCK.

Semiconductor Replacement. Replacement semiconductors should be of the original type or a direct replacement. If the replacement semiconductor is not of the original type, check the manufacturer's basing diagram for proper lead identification.

Free Standing Components. When replacing any components that are free-standing (not directly mounted to circuit boards), be sure to place the new components in the same physical location and position as the old components. If this is not done, there may be a possibility of components touching and causing a short circuit.

4.4 REPACKAGING FOR SHIPMENT

If the Model 252 is to be shipped back to ESI for service or repair, attach a tag showing: owner (with address) and the name of an individual at your firm that can be contacted, complete instrument serial number and a description of the service required.

Save and re-use the package in which your instrument was shipped. If the original packaging is unfit for use or not available, repackage the instrument as follows:

1. Obtain a carton of corrugated cardboard having inside dimensions of not less than six inches more than the instrument dimensions; this will allow for cushioning.

2. Surround the unit with polyethylene sheeting to protect the finish of the instrument.

3. Cushion the instrument on all sides by tightly packing dunnage of urethane foam between the carton and the instrument allowing three inches on all sides.

4. Seal the carton with shipping tape or an industrial stapler.

5.2 MOTHERBOARD CIRCUIT ASSEMBLY (Part No. 43064)

CIRCUIT NO.	DESCRIPTION	PART NO.
C1, C4	Capacitor, 0.047µF, 200V, Mylar	12239
C2, C5, C13 C17, C18, C22, C29	Capacitor, 1µF, 35V, Tant	06472
C3, C37, C39	Capacitor, 1000pF, 500V, Poly	07094
C6	Capacitor, 3.3pF, 1kV, Disc	01925
C7, C12	Capacitor, 6.8µF, 35V, Tant	25339
C8	Capacitor, 2000pF, 500V Mica	02158
C9, C19, C23, C27, C28	Capacitor, 0.1µF, 200V, Mylar	12121
C10, C11, C14, C15, C31, C38, C40	Capacitor, 2.2µF, 20V, Tant	13283
C16	Capacitor, 0.33µF, 200V, Mylar	12446
C20, C21	Capacitor, 0.0159µF, 1%, 100V	18845
C24-C26, C42-C45	Capacitor, 0.05µF, 50V, Disc	12116
C30	Capacitor, 390pF, 600V, Poly	29299
C33	Capacitor, 0.01µF, 100V, Mylar	12260
C34	Capacitor, 0.0033µF, 100V, Mylar	20879
C35	Capacitor, 91pF, 2%, 600V Silver Mica	43738
C36	Capacitor, 0.0022µF, 100V, Mylar	12399
C41	Capacitor, 220pF, 600V, Poly	29297
C46	Capacitor, 560pF, 600V, Poly	25922
CR1, CR2, CR4, CR5 CR9-CR12, CR15	Diode, 1N914A	12356
CR3	Diode Bridge, Mini DIP	26339
CR6-CR8	Diode, 1N4005	01779
CR13, CR14	Diode, 1N4732, Zener	12449
CR16, CR17	Diode, 1N4738, Zener, 8.2V	12160
CR18	Transorb, 10V	42632
J1	Connector, 24 pin, PC	43081
K1	Relay, 12V, 4PDT, PC MT	43083
Q1	Transistor, 2N3702	12041
R1-R3, R8, R29, R32, R41, R46, R56, R87	Resistor, 1kΩ, 10%, 1/4W	13920
R4, R5, R9, R10 R18, R68, R70	Resistor, 100kΩ, 10%, 1/4W	13945
R6	Resistor, Trimmer, 200Ω, 15 Turn	12083
R7	Resistor, 30.1kΩ, 1%, 1/8W	43082
R11-R17	Resistor, Trimmer, 100kΩ, 1 Turn	43092
R19	Resistor, 18kΩ, 10%, 1/4W	13936
R20, R28	Resistor, 2.2kΩ, 10%, 1/4W	13924
R21	Resistor, 220Ω, 10%, 1/4W	13911
R22	Resistor, 1.001Ω, ESI, QB, 4T	43302
R23, R25	Resistor, 1kΩ, ESI, QB	24940
R24	Resistor, 100Ω, 0.02%, ESI, QB	24939
R26	Resistor, 10kΩ, ESI, QB	24941
R27	Resistor, 100kΩ, 0.1%, MF	43300
R30	Resistor, 1.8kΩ, 10%, 1/4W	13923

SECTION 5

PARTS LISTS AND DIAGRAMS

5.1 MAINFRAME PARTS (Part No. 30252)

DESCRIPTION	PART NO.	SEE FIGURE NO.
Assembly, Back Panel (includes	43157	5-4
Circuit Assembly, Power Supply)	(42903)	(5-5)
Assembly, Cable, KELVIN KLIPS®	43072	
Assembly, Range Switch	43151	
P1 Header, 16 Pin, DIP	20799	
S1 Switch, Range	42973	
Circuit Assembly, Display	42916	5-3
Circuit Assembly, Motherboard	43064	5-1,5-2
Cover, Chasis	42988	
Display Window, Acrylic	43116	
Foot, Rubber	08739	
Frame, Front Panel	43068	
Front Panel Overlay	42961	
Handle, Kit	43088	
Instruction Manual	43158	
Knob, Range Switch	22593	
Pushrod, Power Switch	43143	

Options:

Front Panel Dust Cover	43374
Sorting Fixture	Model 2001
Sorting Fixture Cable Assembly (4 BNC)	43586

CIRCUIT NO.	DESCRIPTION	PART NO.
R31, R35, R65	Resistor, Trimmer, 1kΩ, 15 Turn	18469
R33	Resistor, 33Ω, 10%, 1/4W	43301
R34	Resistor, 3.32kΩ, 1%, 1/8W	21771
R36-R40, R79	Resistor, 10kΩ, 10%, 1/4W	13933
R42, R47	Resistor, 68Ω, 10%, 1/4W	13902
R43, R45, R50, R85	Resistor, Trimmer, 10kΩ, 15 Turn	20145
R44	Resistor, 11kΩ, 1%, 1/4W	13359
R48, R49	Resistor, 90.9kΩ, 1%, 1/4W	24557
R51, R58	Resistor, 100kΩ, 1%, 1/4W	21750
R52	Resistor, Trimmer, 2kΩ, 15 Turn	12084
R53	Resistor, 2.2MΩ, 10%, 1/4W	13962
R54, R62	Resistor, Trimmer, 100kΩ, 15 Turn	21455
R55, R86	Resistor, 470Ω, 10%, 1/4W	13915
R57	Resistor, Trimmer, 20kΩ, 15 Turn	21199
R59, R75, R76	Resistor, 22kΩ, 10%, 1/4W	13937
R60, R84	Resistor, 4.99kΩ, 1%, 1/4W	21737
R61, R77	Resistor, 470kΩ, 10%, 1/4W	13955
R63, R64	Resistor, 4.7kΩ, 10%, 1/4W	13927
R66, R80	Resistor, 2kΩ, 1%, 1/4W	21733
R67	Resistor, 82kΩ, 10%, 1/4W	24812
R69	Resistor, 100Ω, 10%, 1/4W	13907
R71	Resistor, Trimmer, 500Ω, 15 Turn	12093
R72	Resistor, 20kΩ, 1%, 1/4W	21743
R73	Resistor, 31.6kΩ, 1%, 1/4W	21745
R74	Resistor, 63.4kΩ, 1%, 1/4W	21748
R78	Resistor, 316kΩ, 1%, 1/4W	21755
R82	Resistor, 10kΩ, 1%, 1/8W, MF	21740
R83	Resistor, 402kΩ, 1%, 1/8W, MF	21756
S2-S6	Switch, 5 Station, Pushbutton	42882
T2	Transformer, Coupling	42964
U1, U3, U13, U17	IC, 3403	40840
U2	IC, 759	43098
U4	IC, 4051	40841
U5, U9, U20, U24	IC, 4053	20744
U6, U8	Resistor Network, 1kΩ, 0.5%, 8 Resistors	43117
U7, U14, U16, U19, U21, U23, U28	Resistor Network, 10kΩ, 0.5%, 8 Resistors	43077
U10	LED Isolator, CLM 6000	25940
U11, U12	IC, 4052	20743
U15	IC, 339	40849
U18	IC, Precision 2.5V REF, 1403	43099
U22, U30	IC, 741	20668
U25, U36	IC, 356H	41473
U26	IC, 741K	21625
U27	IC, TL074	43299
U29	IC, 4012	43086
U31	IC, 14433	43067

CIRCUIT NO.	DESCRIPTION	PART NO.
U32	IC, 4049	41305
U33	IC, 6331N	43104
U34	Resistor Network, 22kΩ, 13 Res	20712
U35	IC, 3080	20723
	PC Board, Motherboard	43063
	Socket, 8 Pin, DIP for U18, U22, U25, U26, U30, U35 and U36	22410
	Socket, 14 Pin, DIP for J5, U1, U3, U13, U15, U17, U27 and U29	19189
	Socket, 16 Pin, DIP for J4,U4,U5, U9,U11,U12,U20,U24,U32 and U33	20860
	Socket, 24 Pin, DIP for U31	41492


D
G
R
C
L
S6
S5
S4
S3
S2
J1
R2
R1
CR16
R8
R3
C4
C1
CR17
R18
K1
U10
U31
CR15
C25
C26
C24
R28
C28
R78
Q1
CR8
R77
C27
R29
CR5
R9
R4
CR4
CR2
CR1
R22
R23
R24
R26
R25
C5
R10
C2
R5
U34
U33
U32
J5
J4
U29
U13
C37
U9
R82
U12
R27
C35
CR6
CR7
U1
CR10
CR9
C6
U36
R79
R83
C8
C7
U14
U4
R11
R19
R20
R21
CR3
R48
U23
R49
U21
C3
R12
R17
C9
C30
C10
C11
R32
R33
R30
U5
R13
R16
U11
U6
C46
R14
R15
C33
R6
U17
C16
U20
C39
R84
R31
R7
R76
R75
R39
R38
R37
R36
R85
R63
R64
R59
CR11
U15
R54
U16
U3
U8
C31
U7
C12
R40
C29
CR12
R73
R34
R74
R72
R66
C41
C38
C34
C19
C36
R65
R53
U24
U19
R67
R68
R69
R70
R35
U18
R44
C14
C15
C13
R42
R80
R41
C23
R51
T2
U22
U26
C22
R58
U28
U27
C21
C20
U35
C43
C42
R45
U30
U25
R61
R52
R87
CR18
R43
C17
R60
R47
C18
R71
R50
CR14
R46
CR13
R86
R62
U2
R57
C40
R56
R55
C44
C45

Figure 5-1. Motherboard



NOTE:

1. ALL RESISTOR VALUES ARE IN OHMS, 1/4W, 10% UNLESS OTHERWISE STATED.
2. ALL CAPACITORS ARE IN μF UNLESS OTHERWISE STATED.
3. LARGE BOLD ALPHABET LETTERS DESIGNATED ON PC BOARD FOR TRIMPOT REFERENCE.
4. ASTERISK (*) DENOTES EQUIVALENT RESISTANCE VALUES FROM SERIES OR PARALLEL COMBINATIONS WITHIN THE RESISTANCE NETWORK PACKAGES.



Figure 5-2. Motherboard Circuit



NOTE:

5 S2 WHEN ENGAGED, RELEASES C, R, OR G FUNCTIONS (S3, S4 & S5).



Figure 5-2. Motherboard Circuit (Continued)

5.3 DISPLAY CIRCUIT ASSEMBLY (Part No. 42916)

CIRCUIT NO.	DESCRIPTION	PART NO.
CR1	Diode, IN914A	12356
Q1-Q4	Transistor, 2N3702	12041
Q5	Transistor, 2N3704	12077
R1-R4	Resistor, 1kΩ, 10%, 1/4W	13920
R5	Resistor, 22kΩ,10%, 1/4W	13937
R6	Resistor,220Ω, 10%, 1/4W	13911
R7	Resistor, 100Ω, 10%, 1/4W	13907
R8,R10	Resistor, 4.7kΩ, 10%, 1/4W	13927
R9	Resistor, 150Ω, 10%, 1/4W	13909
U1	IC, 4049	41305
U2-U4	IC, 4610, 7 Segment Display LED	41350
U5	IC, 4630, ±1 Display LED	41349
U6	IC, 4052	20743
U7	IC, 74LS47	43078
U8	Resistor Network, 150Ω	43080
	PC Board, Display	42915



Figure 5-3. Display Circuit Assembly

5.4 BACK PANEL ASSEMBLY (Part No. 43157)

CIRCUIT NO.	DESCRIPTION	PART NO.
F1	Fuse, 115V Line Voltage, 1/16A 250VAC Slow Blow	08952
F1	Fuse, 230V Line Voltage, 1/20A 250VAC Slow Blow	53327
S1	Power Switch	23164
U1	Voltage Regulator, 7805	20699
U2	Voltage Regulator, LM320T-5	41876
	Back Panel	42986
	Ckt Assy, Power Supply	42903
	Bracket, Instrument Cord Wrap	43768
	Bracket, Switch Support	43052
	Wrap, Switch Support	43138
	Fuse Holder Body, Low Profile	45967
	Fuse Carrier, 3AG, Gray	45966
	Fuse Carrier, 5x20mm, Black	45965
	Power Cord	24077



Figure 5-4. Rear Panel Assembly

5.5 POWER SUPPLY CIRCUIT ASSEMBLY (Part No. 42903)

CIRCUIT NO.	DESCRIPTION	PART NO.
C1, C2	Capacitor, 500µF, 25V	42938
C3, C4	Capacitor, 1µF, Tantalum	06472
C5, C6	Capacitor, 0.05µF, Disc	12116
C7, C8	Capacitor, 100µF, 12V	06157
CR1	Bridge, Full Wave	21236
J3	Connector, 6 Position	43101
S2	Switch, Slide (230/115V)	29233
T1	Transformer	43303
TB1	Barrier Strip, 5 Position	43079
W1	Cable Assembly, Chassis Ground	44699
	PC Board, Power Supply	42902
	Shorting Strap	43308



Figure 5-5. Power Supply

THE GLOBAL SOURCE FOR PROVEN TEST & MEASUREMENT TECHNOLOGY

TEGAM

TEN TEGAM WAY · GENEVA, OHIO 44041
PHONE 216-466-6100 · FAX 216-466-6110

WARRANTY OF TRACEABILITY

The reference standards of measurement of Tegam, Inc., are compared with the U.S. National Standards through frequent tests by the U.S. National Institute of Standards and Technology. The Tegam working standards and testing apparatus used are calibrated against the reference standards in a rigorously maintained program of measurement control.

The manufacture and final calibration of all Tegam instruments are controlled by use of the Tegam reference and working standards and testing apparatus in accordance with established procedures and documented results. (Reference ANSI/NCSL 2540-1-1994)

Final calibration of this instrument was performed with reference to the mean values of the Tegam reference standards or to ratio devices that were verified at the time and place of use.

DISCLAIMER of IMPLIED WARRANTIES

The foregoing warranty of Tegam is in lieu of all other warranties, expressed or implied. Tegam specifically disclaims any implied warranties of merchantability or fitness for a particular purpose. In no event will Tegam be liable for special or consequential damages. Purchaser's sole and exclusive remedy in the event any item fails to comply with the foregoing express warranty of Tegam shall be to return the item to Tegam, shipping charges prepaid, and at the option of Tegam obtain a replacement item or a refund of the purchase price.


95319